



Scale & Flowmeter Specialists

P.O. Box 692, Germantown, WI 53022
Office: (262) 251-3939 Fax: (262) 251-9389

Calibration Summary Report

Date: 1/11/2012

Customer	Michels Corporation	Contact	Scott Wempner
Location	817 West Main Street, Brownsville, WI		Phone 920-539-7785
Make	Badger	Model	4" Prop
Register			Serial: 14965613
Product	Water	Location	Conversion Factors

Unit As Found:

Test	Indicated Quantity (in gallons)	Actual Quantity (in gallons)	Error By Gallons	Error By %
Test # 1	100.00	96.00	-4.00	-4.00%
Test # 2	100.00	97.00	-3.00	-3.00%
Test # 3	100.00	97.30	-2.70	-2.70%

Unit As Left:

Test	Indicated Quantity (in gallons)	Actual Quantity (in gallons)	Error By Gallons	Error By %
Final # 1	100.00	98.50	-1.50	-1.50%
Final # 2	100.00	99.00	-1.00	-1.00%
Final # 3	100.00	99.00	-1.00	-1.00%

Comments Tested meter. Replaced worn propeller and propeller nose bearing.
 Changed gear train.
 Re tested meter after repair.
 Unit passed.

Technician *Bruce Wankowski*

Traceability. Handbook 44, section N.4.1, N.4.2 and N .4.3
N.I.S.T Certificate # W10-104a
State License # 150957-WM
Report prepared in accordance with the written protocol of Accurate Equipment

